

EVOLVE[®] Proline

Radial Head System

SURGICAL TECHNIQUE



Contents

Preface	3	Design Rationale
Chapter 1	5	Preoperative Planning
Chapter 2	6	Surgical Technique
	6	Skin Incision
	7	Direct Lateral Dissection
	8	Resection
	9	Trial Head Selection
	9	Stem Broaching
	10	Neck Planing
	10	Trial Stem Selection
	11	Trial Stem and Head Insertion
	12	Validate Trial Sizing
	12	Trial Head and Stem Removal
	12	Implant Insertion Using Back Table Implant Assembly
	13	Implant Insertion Using <i>In Situ</i> Assembly
	13	Locker Assembly
	14	Implant Locking
	15	Locker Insertion in Very Tight Elbows (optional)
	16	Closure
	17	Post-Op Care
Appendix	18	Ordering Information
	18	EVOLVE® Proline Implants
	19	EVOLVE® Proline Instruments
	19	EVOLVE® Locker Instruments

EVOLVE® Proline Radial Head System

Surgical Technique as described by Graham King, MD

Proper surgical procedures and techniques are the responsibility of the medical professional. The following guidelines are furnished for information purposes only as techniques used by Graham King, MD. Each surgeon must evaluate the appropriateness of the procedures based on his or her personal medical training and experience. Prior to use of the system, the surgeon should refer to the product package insert for complete warnings, precautions, indications, contraindications and adverse effects. Package inserts are also available by contacting Wright Medical Technology, Inc.

Design Rationale

Presented by Graham King, MD

The EVOLVE® Proline RH System is the culmination of years of laboratory and clinical research as well as over 16,000¹ clinical implantations of the EVOLVE® Radial Head prosthesis. It is the state of the art for modular radial head arthroplasty.

The smooth stem design continues to be utilized and is now supported by numerous studies²⁻⁵ including two long term studies. The Harrington study,⁴ published in the *Journal of Trauma* in 2001, included a patient cohort of 20 patients with acute comminuted radial head fractures, with a mean follow-up of 12 years with a range of 6-29 years. This study concluded that a smooth stem radial head “functions well on a long-term basis”. It also concluded that “good to excellent results can be anticipated in approximately 75% of patients and the overall complication rates are acceptable.” The King study,⁵ presented at the Annual ASSH Meeting in September, 2006, included a consecutive patient cohort of 32 patients who underwent smooth stem radial head arthroplasty for elbow reconstruction with a mean follow-up of 8 years. The King study concluded that “metallic radial head arthroplasty for elbow reconstruction is a safe and durable procedure that provides patients with long term functional range of motion and pain relief.”

Conceptually, the annular ligament guides the motion of the EVOLVE® Radial Head prosthesis optimally with the capitellum and the proximal radial ulnar joint rather than relying on the motion patterns of the radial neck. Given that the native radial head is not circular and the articulation with the capitellum is usually offset from that of the radial neck, there is a natural cam effect which occurs during forearm rotation that is difficult for an off-the-shelf axisymmetric implant to replicate. Even eccentrically designed prosthetic implants cannot precisely reproduce the native anatomy and motion patterns due to the highly variable shape of the proximal radius.⁶ The EVOLVE® Radial Head prosthesis utilizes a spacer concept with a smooth stem. The smooth stem can move slightly in the proximal radius so that the radial head tracks optimally with the articular surfaces, reducing abnormal kinematics and therefore problems with articular wear and pain. While the slight movement of the smooth stem in the radial neck can cause some radiolucency, this is not a source of concern. Our long term experience with this design shows that this radiolucency does not progress beyond 1 year and is not a source of pain.

(continues)

1. Internal sales data as of November, 2006.
2. Grewal R, MacDermid J, Faber K, Drosdowech D, King, G. Comminuted radial head fractures treated with a modular metallic radial head arthroplasty. *Journal of Bone and Joint Surgery*. October, 2006.
3. Moro JK, Werier J, MacDermid JC, Patterson SD, King GJ. Arthroplasty with a metal radial head for unreconstructible fractures of the radial head. *Journal of Bone and Joint Surgery*. August, 2001.
4. Harrington IJ, Sekyi-Out A, Barrington TW, Evans DC, and Tuli V. The functional outcome with metallic radial head implants in the treatment of instable elbow fractures: a long term review. *Journal of Trauma*. Jan. 2001.
5. Shore B, MacDermid J, Faber K, King G. Outcome of metal radial head arthroplasty in elbow reconstruction. Annual Meeting of ASSH, Sept. 2006.
6. King G, Zarzour Z, Patterson S, Johnson J. An anthropometric study of the radial head. *The Journal of Arthroplasty*. 16:112-116, 2001.



Design Rationale (cont'd)

In contrast, a malarticulating implant with a well-fixed stem causes high contact pressures on the opposing articular cartilage and can lead to early failure.

An alternative approach is the use of a bipolar articulation. While this at first seems attractive, the issues of polyethylene wear and particulate debris are a real concern given the relatively young age at which most radial head implants are employed. Furthermore, a bipolar implant is less effective at maintaining elbow and forearm stability due to a tendency of the articulation to angulate under load.

The EVOLVE® Proline RH System continues to offer the two part, modular implant design that gives surgeons the ability to appropriately match the patient's anatomy. The original sizing of the implant system is based on an anthropometric study⁶ of the proximal radius. This research demonstrated a wide variability in the size and shape of the radial head as well as a poor correlation of the size of the radial head with the dimensions of the medullary canal of the radial neck. Based on this study and extensive clinical experience, the EVOLVE® Proline RH System head sizes now range from 18 to 28mm in diameter and stem sizes now range from 4.5 to 9.5mm diameter (Figure 1). Furthermore, the system now has three head heights and three stem heights that enable precise replication of the native radial head articulation with the proximal radioulnar joint.




Figure 1

Every instrument in the EVOLVE® Proline RH System has been updated to give surgeons a simpler and more precise technique. The broaching, trialing and implant assembly instrumentation has been enhanced. The most dramatic change is the new *in situ* assembly device which allows for easier implant insertion and less surgical trauma to the joint. The *in situ* Locker provides confident implant assembly, even in small, tight elbows with intact ligaments. The Locker engages the reliable Morse taper connection of the EVOLVE® implant; there are no set screws or polyethylene to wear out or fail over time. Our testing has shown that an implant locked with the new *in situ* Locker would require 1195N of actual force on average before disassembly could occur.⁷

7. Internal test data available.

Preoperative Planning



Undisplaced radial head fractures should be managed non-operatively while displaced radial head fractures should be treated with open reduction and internal fixation if technically feasible. For those cases, the EVOLVE® Radial Head Reconstruction System can be used (p/n 4911KIT1/A). Comminuted displaced radial head fractures, which cannot be reconstructed with stable internal fixation, should be managed with radial head excision or prosthetic replacement.

In the setting of an associated elbow dislocation, radial head excision without replacement is contraindicated due to valgus instability arising from concomitant injury to the medial collateral ligament of the elbow. The diagnosis of disruption of the medial collateral ligament and/or interosseous membrane is more problematic in patients without an associated elbow or distal radioulnar joint dislocation. In one study, all patients with comminuted radial head fractures without an associated elbow dislocation had insufficiency of the medial collateral ligament or interosseous membrane as documented by stress radiographs.⁸ Given this high frequency of unrecognized soft tissue injury with comminuted radial head fractures, it is not surprising that some authors recommend that primary prosthetic substitution should be performed in all patients where radial head resection is required.

Dr. Graham King and colleagues at St. Joseph's Health Centre in London, Ontario, Canada, performed a cadaveric study evaluating the ability of radial head implants to stabilize the medial collateral ligament deficient elbow. Their findings showed that metallic implants improved elbow stability as measured by a significant decrease in varus-valgus laxity. Additionally, they found that elbow stability following radial head resection and metallic implant arthroplasty was similar to the stability of an intact radial head in the medial collateral ligament deficient elbow.⁹ Their laboratory has also demonstrated the kinematics of the elbow are altered following radial head excision even with intact ligaments. Radial head replacement with the EVOLVE® system can restore the kinematics of the elbow similar to that with the native radial head.¹⁰ This suggests that routine replacement of the radial head may be beneficial, even in the setting where the ligaments are competent.

EVOLVE® Radial Head Replacement is also valuable for the management of reconstructive elbow problems including radial head non-unions and malunions as well as for revision of failed radial head arthroplasty. It is also useful for treating elbow and forearm instability after radial head resection.

8. Davidson PA, Moseley JB, Tullos HS. Radial head fracture: A potentially complex injury. *Clinical Orthopaedics and Related Research*. 297:224-230, 1993.

9. King GJ, Zarzour ZD, Rath DA, Dunning CE, Patterson SD, Johnson JA. Metallic radial head arthroplasty improves valgus stability of the elbow. *Clinical Orthopaedics and Related Research* 368:114-25, 1999.


10. Beingessner DM, Dunning CE, Gordon KD, Johnson JA, King GJ. The effect of radial head excision and arthroplasty on elbow kinematics and stability. *Journal of Bone and Joint Surgery*. 86A:1730-1739, 2004.

Skin Incision

Radiographs of the contralateral elbow and both wrists are helpful in preoperative planning, particularly if the radial head has previously been excised. Estimate the radial head and stem sizes needed using the Proline X-ray Template (p/n 496XR01).

With the patient in either the supine or lateral decubitus position, make a posterior midline longitudinal skin incision just lateral to the tip of the olecranon. Elevate a full thickness lateral flap (fasciocutaneous) on the deep fascia to protect the cutaneous nerves. The posterior midline incision permits access to the medial side of the elbow if repair of the medial collateral ligament is necessary to restore elbow stability. It is also more cosmetic than a laterally-based incision. In patients with isolated injuries to the radial head, a traditional lateral skin incision may be employed. However, first identify and protect the cutaneous nerves which usually cross the incision (Figure 2).


Figure 2



Direct Lateral Dissection

Pronate the forearm to move the posterior interosseous nerve more distal and medial during the surgical approach. Split the extensor digitorum communis tendon longitudinally at the midaspect of the radial head and incise the underlying radial collateral and annular ligaments (Figure 3). Keep dissection anterior to the lateral ulnar collateral ligament to prevent the development of posterolateral rotatory instability. If additional exposure is needed, elevate the humeral origin of the radial collateral ligament and the overlying extensor muscles anteriorly off the lateral epicondyle and lateral supracondylar ridge. In the unusual circumstance where further exposure is required, consider releasing the posterior component of the lateral collateral ligament (including the lateral ulnar collateral ligament). However, careful ligament repair is required at the end of the procedure in order to restore the varus and posterolateral rotatory stability of the elbow. In many circumstances, the radial head is easily visualised after opening the subcutaneous tissue due to avulsion of the lateral collateral ligament and common extensor muscles from the lateral epicondyle during the injury.

Figure 3



Resection

Remove and retain all fragments of the radial head. Using a sagittal saw, resect the remaining radial head at the level of the radial neck fracture, perpendicular to the neck to make a smooth surface for seating the prosthetic radial head (Figure 4). Confirm complete radial head excision with an image intensifier and by reassembling the resected radial head in the Sizing and Assembly Dish (p/n 24981005) (Figure 5). It is recommended that at least 60% of the native radial neck be in contact with the implant. If not, make the radial neck cut more distal and use a thicker head/stem prosthesis. Copiously irrigate the joint to remove all loose intra-articular debris. Evaluate the capitellum for chondral injuries or osteochondral fractures. Manage associated fractures of the coronoid as indicated prior to radial head replacement. Carefully place a Hohman retractor around the posterior aspect of the proximal radial neck to deliver the radial neck laterally (Figure 6). Avoid placing the retractor anteriorly due to the risk of injury to the posterior interosseous nerve from pressure.




Figure 4




Figure 5




Figure 6




Figure 7



Figure 8



Figure 9

Trial Head Selection

Select the appropriate Trial Head (p/n 2499Hxxx) diameter based on backtable reassembly of the radial head fragments. For elliptically shaped radial heads, select the minimum rather than the maximum diameter (Figure 7). Pay special attention to replicate the size of the articular dish rather than the outside diameter of the native head (Figure 8). Select the prosthesis height based on the thickness of the flatter articular portion of the native radial head that articulates with the proximal radial ulnar joint (Figure 9). In the setting where the radial head has been previously excised, use the Proline X-ray Template on the contralateral normal radial head to determine the appropriate diameter and height of the radial head implant. If the native radial head is between available implant sizes in diameter or height, downsize the implant in the appropriate dimension.

Stem Broaching

Create an opening in the medullary canal using the Starter Awl (p/n 24987100). Sequentially ream the radial neck by inserting the Stem Broaches (p/n 24987xxx) to the depth indicators on the Broaches (Figure 9) until the Stem Broaches no longer pass easily into the canal due to cortical contact.




Figure 10




Figure 11a




Figure 11b

Neck Planing

Leave the last stem broach in the canal and remove the handle. Slip the Neck Planer (p/n 24981003) over the Stem Broach (Figure 11a). Gently rotate the Neck Planer to create a smooth contact surface on the radial neck, perpendicular to the longitudinal axis of the radial neck (Figure 11b). Avoid excessive planing as it may increase the height of the stem required.

Trial Stem Selection

Select the appropriate Trial Stem (p/n 2499Sxxx) diameter based on the largest Stem Broach that easily fits in the canal. The Trial Stem should fit into the radial neck (Figure 12) without force and have a slightly loose but not sloppy fit in the medullary canal of the radius. **Undersizing the Trial Stem diameter by one size is recommended in most cases to allow for the implant to toggle and precisely conform with the capitellum during range of motion.** Select the stem collar height by placing the trial stem into the trial head and comparing the total height with that of the native radial head that was excised (Figure 13).



Figure 12




Figure 13

Trial Stem and Head Insertion

Grasp the Trial Stem with the Trial Stem Handle (p/n 24981002) so that the handle sits below the Trial Head. Insert the Trial Stem into the medullary canal. Screw the Trial Head onto the Trial Head Handle (p/n 24981001). Holding the Trial Head Handle in line with the Trial Stem Handle, slide the Trial Head over the Trial Stem platform (Figure 14a). Once the Trial Head is completely seated on the Trial Stem platform, rotate the Trial Handles 90° apart (Figure 14b) to lock the Trial Head and Trial Stem together via a ball plunger connection (Figure 14c). If the Trial Handles do not rotate easily, reconfirm that the Trial Head is completely seated on the Trial Stem platform.




Figure 14a




Figure 14b




Figure 14c

Validate Trial Sizing

Unscrew the Trial Head Handle from the Trial Head and remove the Trial Stem Handle. Reduce the elbow with the trials in place. Verify smooth motion in passive flexion and extension of the elbow and rotation of the forearm. Some translation of the Trial Head relative to the capitellum is normal with forearm rotation. Assess the appropriate implant height by pronating the forearm to compensate for the lateral destabilization induced by the surgical approach or injury. The Trial Head should articulate with the most proximal margin of the proximal radioulnar joint approximately 1 mm distal to the coronoid process.

NOTE: To reduce the risk of cartilage wear on the capitellum from excessive pressure, avoid overstuffing the radiocapitellar joint with a radial head implant that is too thick. To avoid overstuffing the radial-capitellar joint, use the combined Trial Head and Trial Stem collar height to approximate the height of the native radial head and radial neck portion that was resected, not the gap between the radial neck and the capitellum. There is often a small gap between the Trial Head and Capitellum; particularly in cases with lateral ligament injuries. Do not increase the implant thickness to compensate for the ligament injuries. Repairing the collateral ligaments prior to closure will stabilize the joint.



Figure 15

Use an image intensifier to evaluate ulnar variance at the wrist. An implant that is too thick will have ulnar negative variance and an implant that is too thin will have ulnar positive variance relative to the contralateral wrist. Visualize the medial ulnohumeral joint in an anteroposterior view with an image intensifier to ensure that the joint space is symmetrical (Figure 15). An implant that is too thick will result in varus alignment and a non-parallel medial ulnohumeral joint space that is wider laterally. If the prosthesis is tracking poorly on the capitellum with forearm rotation, trial a smaller stem size to ensure that the articulation of the radial head with the capitellum is controlled by the annular ligament and articular congruency, and not dictated by the motion pathways of the proximal radial shaft.

NOTE: A metallic radial head will appear larger on x-ray than the native radial head because it is replacing radiolucent cartilage as well as radiographic bone.

Trial Head and Stem Removal

Once optimal sizing has been determined, reattach the Trial Handles to the Trials. Unlock the Trial Head from the Trial Stem by realigning the handles. Remove the Trial Head from the joint space and then remove the Trial Stem. Irrigate the joint thoroughly.

Implant Insertion Using Back Table Implant Assembly

In most acute injuries, the proximal radius is sufficiently mobile or the lateral ligaments have been compromised such that the implant can be assembled on the back table and inserted as a monoblock implant. To do this, insert the Stem Implant (p/n 496Sxxx) into the Head Implant (p/n 496Hxxx) and place onto the Sizing and Assembly Dish. Place the appropriately sized Stem Impactor (p/n 24981007-24981009) over the stem and strike it firmly three times with a mallet (Figure 16). Insert the assembled implant into the proximal radius by retracting the proximal radius laterally (Figure 6).



Figure 16



Figure 17a



Figure 17b

Implant Insertion Using *In Situ* Assembly

When the lateral ligaments are intact in acute injuries and in cases of late reconstruction, insertion of the assembled implant may not be possible due to insufficient mobility of the proximal radius. In these settings, the two components of the implant should be inserted separately and then coupled *in situ* using the supplied locker.

While retracting the proximal radius with a retractor, insert the Stem Implant into the medullary canal. It should slide in easily (Figure 17a). Using finger control, slide the Head Implant into the joint space with the Head Implant female taper over the Stem Implant male taper (Figure 17b).

Locker Assembly

Assemble the Locker by first inserting the Stem Paddle Post (p/n 24991001) into the Locker Body (p/n 24991000). Screw the Locker Assembly Knob (p/n 24982005) onto the Stem Paddle Post. Insert the appropriately sized Head Paddle (p/n 24991018-24991028) into the jaw on the Locker Body. Insert the appropriately sized Stem Paddle (p/n 24991045-24991095) into the jaw on the Stem Paddle Post (Figure 18).

NOTE: The Locker is the only recommended device for *in situ* assembly. A tamp and/or mallet will not generate enough force to adequately secure the Morse taper and disassociation may occur.

LOCKER ASSEMBLY

1. Insert Stem Paddle Post (A) into Locker Body (B)
2. Screw Locker Assembly Knob (C) onto Stem Paddle Post (A)
3. Snap Head Paddle (D) into Locker Body (B)
4. Snap Stem Paddle (E) into Stem Paddle Post (A)




Figure 18

Implant Locking

With traction on the arm, gently slide the Stem and Head Paddles into the joint space to avoid damaging the capitellum. Once the Locker is properly seated on the implant (Figure 19a), tighten the Locker Assembly Knob and give the Locker one firm squeeze (Figure 19b). Unscrew the Locker Assembly Knob to disengage the Locker from the now assembled implant.

Note: Because of the tremendous load being applied by the In Situ Locker, on some occasions, after assembling the implant, the locker jaws will not release freely. In those cases, loosen the Locker Assembly Knob 2-3 turns and lightly tap the end of the Locker Assembly Knob with a small mallet.



Figure 19a




Figure 19b

Locker Insertion in Very Tight Elbows (optional)

In some cases, the elbow joint may be too small or tight to allow both the Head Paddle and Stem Paddle to be inserted concurrently. In those situations, a consecutive approach can be used. Instead of snapping the Stem Paddle into the Stem Paddle Post (Figure 18, step 4), use the Trial Head Handle to hold onto the Stem Paddle (Figure 20a). Insert the Stem Paddle underneath the Stem Implant collar. Carefully guide the Head Paddle, attached to the Locker Body, onto the Head Implant while also guiding the Stem Paddle Post onto the Stem Paddle (Figure 20b). Once the Locker is positioned correctly, tighten the Locker Assembly Knob and give the Locker one firm squeeze.




Figure 20a




Figure 20b


Closure

Following radial head replacement, repair the lateral collateral ligament and extensor muscle origins back to the lateral condyle. If the posterior half of the lateral collateral ligament is still attached to the lateral epicondyle, repair the anterior half of the lateral collateral ligament (the annular ligament and radial collateral ligament) and extensor muscles to the posterior half using interrupted absorbable sutures (Figure 21). If the lateral collateral ligament and extensor origin have been completely detached either by the injury or surgical exposure, securely repair them to the lateral epicondyle using drill holes through bone and non-absorbable sutures or suture anchors. Place a single drill hole at the axis of motion (the centre of the arc of curvature of the capitellum) and two drill holes placed anterior and posterior to the lateral supracondylar ridge. Employ a locking (Krackow) suture technique to gain a secure hold of the lateral collateral ligament and common extensor muscle fascia. Pull the ligament sutures into the holes drilled in the distal humerus using suture retrievers. Pronate the forearm and avoid varus forces while tensioning the sutures prior to tying. Leave the knots anterior or posterior to the lateral supracondylar ridge to avoid prominence.

Following replacement arthroplasty and lateral soft tissue closure, place the elbow through an arc of flexion-extension while carefully evaluating for elbow stability in pronation, neutral, and supination. Pronation is generally beneficial if the lateral ligaments are deficient, supination if the medial ligaments are deficient and neutral position if both sides have been injured.

In patients who have an associated elbow dislocation, perform additional repair of the medial collateral ligament and flexor pronator origin if the elbow subluxates at 40° or more of flexion. After tourniquet deflation and secure hemostasis, the subcutaneous tissues and skin are closed in layers.

Figure 21



Post-Op Care

The recommended Post-Op Care varies primarily according to ligament competency.^{11,12}



Lateral view at 2 years post-op



AP view at 2 years post-op

1. MCL and LCL Competent

- Splint elbow in extension and forearm in neutral
- Unrestricted active elbow motion permitted postoperatively
- Night-time resting extension splint may assist in gaining terminal extension

2. MCL Competent but LCL Incompetent

- Splint elbow at 90° with forearm pronated
- Active flexion-extension performed with forearm pronated
- Prosupination performed with elbow in flexion
- Avoid extension in supination for six weeks

3. MCL Incompetent and LCL Competent

- Splint elbow at 90° with forearm supinated
- Active flexion-extension performed with forearm supinated
- Prosupination performed with elbow in flexion
- Avoid extension in pronation for six weeks

4. MCL and LCL Incompetent

- Splint elbow at 90° with forearm in neutral
- Active flexion-extension performed with forearm in neutral rotation
- Prosupination performed with elbow in flexion
- Gradually allow increasing extension as stability improves with healing over six weeks

5. General Rehabilitation

- No passive stretching for six weeks to avoid heterotopic ossification.
- Strengthening exercises commence six to eight weeks postoperatively
- Night-time extension splint may be useful to regain terminal elbow extension.
- Prescribing indomethacin may reduce the incidence of heterotopic bone formation.

11. Armstrong AD, Dunning CE, Faber KJ, Duck TR, Johnson JA, King GJW: Rehabilitation of the medial collateral ligament-deficient elbow: An *in vitro* biomechanical study. *J Hand Surg* 25A:1051-1057, 2000.
12. Dunning CE, Zarzour ADS, Patterson SD, Johnson JA, King GJW: Muscle forces and pronation stabilize the lateral ligament deficient elbow. *Clin Orthop & Related Research* 388:118-124, 2001.

Ordering Information

EVOLVE® Proline Implants


2499KITA

Item #	Description	Kit Qty
496H018	HEAD 18MM	1
496H218	HEAD 18MM +2	1
496H418	HEAD 18MM +4	1
496H020	HEAD 20MM	1
496H220	HEAD 20MM +2	1
496H420	HEAD 20MM +4	1
496H022	HEAD 22MM	1
496H222	HEAD 22MM +2	1
496H422	HEAD 22MM +4	1
496H024	HEAD 24MM	1
496H224	HEAD 24MM +2	1
496H424	HEAD 24MM +4	1
496H026	HEAD 26MM	1
496H226	HEAD 26MM +2	1
496H426	HEAD 26MM +4	1
496H028	HEAD 28MM	1
496H228	HEAD 28MM +2	1
496H428	HEAD 28MM +4	1

Item #	Description	Kit Qty
496S045	STEM 4.5MM	1
496S245	STEM 4.5MM +2	1
496S445	STEM 4.5MM +4	1
496S055	STEM 5.5MM	1
496S255	STEM 5.5MM +2	1
496S455	STEM 5.5MM +4	1
496S065	STEM 6.5MM	1
496S265	STEM 6.5MM +2	1
496S465	STEM 6.5MM +4	1
496S075	STEM 7.5MM	1
496S275	STEM 7.5MM +2	1
496S475	STEM 7.5MM +4	1
496S085	STEM 8.5MM	1
496S285	STEM 8.5MM +2	1
496S485	STEM 8.5MM +4	1
496S095	STEM 9.5MM	1
496S295	STEM 9.5MM +2	1
496S495	STEM 9.5MM +4	1



Proline Instrument Tray



Locker Instrument Tray

EVOLVE® Proline Instruments

2499KIT1

Item #	Description	Kit Qty
44112009	AO DRIVER HANDLE	2
24981007	IMPACTOR 4.5/5.5MM	1
24981008	IMPACTOR 6.5/7.5MM	1
24981009	IMPACTOR 8.5/9.5MM	1
24981003	NECK PLANER	1
24981005	SIZING & ASSEMBLY DISH	1
24987100	STEM STARTER AWL	1
24987145	STEM BROACH 4.5MM	1
24987155	STEM BROACH 5.5MM	1
24987165	STEM BROACH 6.5MM	1
24987175	STEM BROACH 7.5MM	1
24987185	STEM BROACH 8.5MM	1
24987195	STEM BROACH 9.5MM	1
24987105	STEM BROACH 10.5MM	1
24981001	TRIAL HEAD HANDLE	1
24981002	TRIAL STEM HANDLE	1
496XR01	PROLINE X-RAY TEMPLATE	1
24981010	INSTRUMENT TRAY	1
2499H018	TRIAL HEAD 18MM	1
2499H218	TRIAL HEAD 18MM +2	1
2499H418	TRIAL HEAD 18MM +4	1
2499H020	TRIAL HEAD 20MM	1
2499H220	TRIAL HEAD 20MM +2	1
2499H420	TRIAL HEAD 20MM +4	1
2499H022	TRIAL HEAD 22MM	1
2499H222	TRIAL HEAD 22MM +2	1
2499H422	TRIAL HEAD 22MM +4	1
2499H024	TRIAL HEAD 24MM	1
2499H224	TRIAL HEAD 24MM +2	1
2499H424	TRIAL HEAD 24MM +4	1
2499H026	TRIAL HEAD 26MM	1
2499H226	TRIAL HEAD 26MM +2	1
2499H426	TRIAL HEAD 26MM +4	1
2499H028	TRIAL HEAD 28MM	1
2499H228	TRIAL HEAD 28MM +2	1
2499H428	TRIAL HEAD 28MM +4	1
24981011	PROLINE REPLACEMENT LID	0

Item #	Description	Kit Qty
2499S045	TRIAL STEM 4.5MM	1
2499S245	TRIAL STEM 4.5MM +2	1
2499S445	TRIAL STEM 4.5MM +4	1
2499S055	TRIAL STEM 5.5MM	1
2499S255	TRIAL STEM 5.5MM +2	1
2499S455	TRIAL STEM 5.5MM +4	1
2499S065	TRIAL STEM 6.5MM	1
2499S265	TRIAL STEM 6.5MM +2	1
2499S465	TRIAL STEM 6.5MM +4	1
2499S075	TRIAL STEM 7.5MM	1
2499S275	TRIAL STEM 7.5MM +2	1
2499S475	TRIAL STEM 7.5MM +4	1
2499S085	TRIAL STEM 8.5MM	1
2499S285	TRIAL STEM 8.5MM +2	1
2499S485	TRIAL STEM 8.5MM +4	1
2499S095	TRIAL STEM 9.5MM	1
2499S295	TRIAL STEM 9.5MM +2	1
2499S495	TRIAL STEM 9.5MM +4	1

EVOLVE® Locker Instruments

2499KIT2

Item #	Description	Kit Qty
24982005	LOCKER ASSEMB KNOB	1
24991000	LOCKER BODY	1
24981012	LOCKER TRAY	1
24991001	STEM PADDLE POST	1
24991045	STEM PADDLE 4.5MM	1
24991055	STEM PADDLE 5.5MM	1
24991065	STEM PADDLE 6.5MM	1
24991075	STEM PADDLE 7.5MM	1
24991085	STEM PADDLE 8.5MM	1
24991095	STEM PADDLE 9.5MM	1
24991018	HEAD PADDLE 18MM	1
24991020	HEAD PADDLE 20MM	1
24991022	HEAD PADDLE 22MM	1
24991024	HEAD PADDLE 24MM	1
24991026	HEAD PADDLE 26MM	1
24991028	HEAD PADDLE 28MM	1
24981013	LOCKER REPLACEMENT LID	0

The EVOLVE® Family of Radial Head Products



EVOLVE® Proline System

18 head sizes and 18 stem sizes
2499KIT1/A



EVOLVE® System

15 head sizes and 10 stem sizes
2497KIT3/2496KITA



EVOLVE® Locker

for use with EVOLVE® Proline System or EVOLVE® System
2499KIT2



EVOLVE® Radial Head Reconstruction

with radial head plate, headed and headless screw options
4911KIT1/A



For Bone Voids, Use

ALLOMATRIX® DR

Peri-articular graft
86DR-0300 3cc



For Ligament Reinforcement, Use

GRAFTJACKET® Matrix – Maximum Force

Regenerative Tissue Matrix
8600-4X07 4x7cm Non-meshed



Wright Medical Technology, Inc.
1023 Cherry Road
Memphis, TN 38117
800 238 7117
901 867 9971
www.wmt.com

Wright Medical EMEA
Atlas Arena, Australia Building
Hoogoorddreef 7
1101 BA Amsterdam
the Netherlands
011 31 20 565 9060

Wright Medical UK Ltd.
Unit 1, Campus Five
Letchworth Garden City
Hertfordshire SG6 2JF
United Kingdom
011 44 (0)845 833 4435